

**Advice 13-2015 of the Scientific Committee of the FASFC on a request for approval of analysis devices used in the official determination of the composition and the quality of raw milk supplied to buyers.**

The Scientific Committee is asked to give an advice on three milk analyzers, for the official analysis of the composition and the quality of raw milk provided to buyers. The Scientific Committee is of the opinion that the performances of the three devices evaluated in accordance with the phase I of the ISO 13366-2 standard and the ISO 9226 /IDF 141C:2000 standard by ACTALIA Cecalait, the AIA, the CRAW and the ILVO are satisfactory. The analysis devices can be certified subject to the completion of Phase II of validation. This signifies after that has been demonstrated by a Belgian laboratory, that this equipment really meets the required limits under routine circumstances.

The full text is available on this website in dutch and in french, respectively under the section "Wetenschappelijk Comité/Adviezen" and "Comité scientifique/Avis".